



Immunophenotyping Control

CD-Chex Plus® is a positive procedural control for monitoring immunophenotyping by flow cytometry. It provides the most assayed CD markers in the industry on a single control, including those on the HIV panel of markers recommended by the CDC and a normal level of CD34+ cells. CD-Chex Plus is manufactured from normal human peripheral blood leukocytes and erythrocytes.

Compatibility

BD Biosciences and Beckman Coulter® flow cytometry systems

Accuracy

CD-Chex Plus verifies the accuracy and reproducibility of the steps involved in the process of the immunophenotyping procedure including:

- Verification of mAb reagents
- Verification of red blood cell lysis
- Validation of instrument performance
- Validation of operator technique

Stability (Store 2 °C - 10 °C)

| | |
|-----------------------|---------|
| Closed-vial stability | 90 days |
| Open-vial stability | 30 days |

Features and Benefits

- Positive procedural control assayed for flow cytometry systems
- Provides the most assayed CD markers in the industry, including T-Lymphocytes, B-Lymphocytes, granulocytes, monocytes and NK cells
- Available in two clinically relevant levels of CD4+ cells

STATS®

Interlaboratory Quality Control program features online accessibility with enhanced real-time reporting and improved account management capability. Open to all customers at no charge, the Streck *STATS* program includes personalized, easy-to-read reports that detail performance, identify trends and facilitate real-time prompt corrective action. More information can be found at www.streck.com.

Ordering Information*

Cap-Pierceable Plastic Vials

| Level | Description | Catalog # |
|-----------------|-------------|-----------|
| Normal | 2 x 3.0 mL | 213365 |
| CD4 Low | 2 x 3.0 mL | 213366 |
| CD4 Low, Normal | 2 x 3.0 mL | 213391 |
| CD4 Low, Normal | 4 x 3.0 mL | 213369 |
| CD4 Low, Normal | 10 x 3.0 mL | 213370 |

Glass Vials

| Level | Description | Catalog # |
|-----------------|-------------|-----------|
| Normal | 2 x 2.5 mL | 213326 |
| CD4 Low | 2 x 2.5 mL | 213331 |
| CD4 Low, Normal | 2 x 2.5 mL | 213340 |
| CD4 Low, Normal | 4 x 2.5 mL | 213341 |
| CD4 Low, Normal | 10 x 2.5 mL | 213344 |

*Other kit configurations available.



The brand and product names of the instruments are trademarks of their respective holders.

800.843.0912 | STRECK.COM 320006-14 2017-07



www.rafer.es