



		FrameStar® 2-Component PCR Plates								FrameStar® PCR Plates, RA Plate, FrameStrips®								Standard PCR Plates and Strips																
Skirt		skirted				semi-skirted				non skirted		Break-A-Way		Break-2-Ways		skirted	strips	skirted		semi-skirted		non	strips				strips							
Number of wells		384	384	96	192	96	96	96	96	96	96	96	96	96	96	96	8	384	384	96	96	96	96	8	8	8	8	8	4					
Standard/Low Profile		/	/	L	/	L	L	S	S	S	S	L	S	L	L	S	/	/	L	S	L	S	S	S	L	S	L	S	/					
Details see PCR Consumables Overview Brochure page		8	8	9	8	9	10	10	10	11	11	11	13	13	13	13	15	16	18	18	18	19	19	19	20/22	20	17	21	21	21				
Product code <sup>1</sup> /Short description		4ti-0384 FrameStar® 384	4ti-0380 FrameStar® 384 Roche	4ti-0960 FrameStar® 96	4ti-0192 FrameStar® 192	4ti-0950 & 4ti-0954 FrameStar® 96 Roche	4ti-0910 FrameStar® 96 FastPlate	4ti-0730 FrameStar® 96 ABI®	4ti-0770 FrameStar® 96 ABI®	4ti-0900 FrameStar® 96	4ti-0710 FrameStar® 96	4ti-0720 FrameStar® 96	4ti-1000 FrameStar® Break-A-Way	4ti-1200 FrameStar® Break-A-Way	4ti-1300 FrameStar® Break-2-Ways	4ti-1400 FrameStar® Break-2-Ways	4ti-0960/RA Random Access 96	4ti-0785 FrameStrip® 8 Well	4ti-1384 384 Well	4ti-1381 384 Well Roche	4ti-0740 96 Well	4ti-0760 96 Well	4ti-0955 96 Well Roche	4ti-0735 96 Well ABI®	4ti-0750 - 96 Well 4ti-0750/TA- dividable	4ti-0781 8 Well	4ti-0753 Vari-Strips™ 8 Well <sup>2</sup>	4ti-0792 8 Well, Attached Flat Caps	4ti-0793 8 Well, Attached Flat Caps	4ti-0794 8 Well, Attached Domed C.	4ti-0796 4 Well Rotor-Gene®			
<b>BIO-RAD MJ RESEARCH continued</b>																																		
qPCR Cyclers	Strips only			4																														
	96 well block			4					✓	✓	✓	✓	✓	✓	✓	✓	✓				✓	✓			✓	✓	4	✓	✓	✓				
				4																														
<b>CORBETT RESEARCH</b>																																		
Thermal Cyclers	96 well block								4	✓																								
	384 well block	4			•														✓															
qPCR Cycl.	Strips only																															4		
<b>EPPENDORF</b>																																		
Thermal Cyclers	96 well block			4					✓	✓											✓	✓			✓	✓	✓	✓	✓	✓	✓			
	384 well block	4			•							4	✓	✓	✓										✓	✓	✓	✓	✓	✓	✓			
qPCR Cycl.	96 well block			4					✓	✓											✓	✓			✓	✓	✓	✓	✓	✓	✓			
<b>GE HEALTHCARE / AMERSHAM</b>																																		
Sequencers	96 well block			4																														
	384 well block	4			•														✓															
<b>PEQLAB / VWR</b>																																		
Thermal Cyclers	Strips only																																	
	96 well block			4					✓	✓																								
	384 well block	4			✓														✓		✓	✓			✓	✓	✓	✓	✓	✓	✓			
<b>ROCHE</b>																																		
qPCR Cyclers	96 well block					4																✓					✓							
	384 well block		4																		✓													
	Strips only																										4							
<b>SENSOQUEST</b>																																		
Thermal Cyclers	96 well block			✓			✓	✓	✓	✓											✓	✓		✓	✓	✓	✓	✓	✓	✓	✓			
	192 well block				✓																													
	384 well block	4			•																													
<b>TAKARA</b>																																		
Thermal Cycl.	96 well block								4	✓																								
<b>TECHNE</b>																																		
Thermal Cyclers	Strips only											4		✓																				
	96 well block								4	✓				✓																				
	384 well block	4			•																													
qPCR Cycl.	96 well block	4		4	✓																													

<sup>1</sup> Short product code shown only (without details on frame and well colour), please refer to the corresponding product page for ordering details of all variations available.  
<sup>2</sup> For compatibility information of the Vari-Plates™ please refer to the table entry for the respective Vari-Plate™ frame.  
 4 Recommended option   ✓ Compatible   • Should be compatible, please check with your specific instrument/block   TA Compatible with separated Tear-A-Way™ Plates, only, please refer to page 22.

The individual instrument-plate compatibility information is for guidance only. Samples of all plates are available before purchase to ensure compatibility. Please check installed heat block, refer to instruments manual for details or contact our technical support.  
 4titude® recognises that designated trademarks and brands of the Instrument Compatibility Table are the property of their respective owners.